



Certificate of Analysis

Product Name: HBsAg Rapid Test Cassette (Serum/ Plasma)	
Catalog No.: IHBSG-C31	
Product Lot No.: HBSG23020002	Batch Quantity: 4000PCS
Mfg Date: 2023.02.07	Expiry Date: 2025.02.06
Strip Size (mm): 60*4	Inspection Date: 2023.04.23
Buffer Lot. /Exp.: /	Remark: /

QC Items		Specification	Test Result	Conclusion
Physical	Appearance	Intact	Intact	PASS
Functional Performance	Positive Sample	100% Positive	100% Positive	PASS
	Negative Sample	100% Negative	100% Negative	PASS

Conclusion: This batch of product meets the QC specifications.

QC: Arno Wang
 Arno Wang

Title: Quality Control Supervisor

Date: 2023.04.23





Certificate of Analysis

Product Name: HBsAg Rapid Test Strip (Whole blood /Serum /Plasma)	
Catalog No.: IHBSG-S41	
Product Lot No.:	Batch Quantity: 4000PCS
Mfg Date:	Expiry Date:
Strip Size (mm): 80*4	Inspection Date: 2023.04.23
Buffer Lot. /Exp.:	Remark: /

QC Items		Specification	Test Result	Conclusion
Physical	Appearance	Intact	Intact	PASS
Functional Performance	Positive Sample	100% Positive	100% Positive	PASS
	Negative Sample	100% Negative	100% Negative	PASS

Conclusion: This batch of product meets the QC specifications.

Arno Wang

QC:

Arno Wang

Title: Quality Control Supervisor

Date:

2023.04.23

